



ALL THREADED FASTENERS TO BE 1/2"-13 GRADE 5, PLATED, TIGHTENED TO 40 ft-lbs.

THIS DRAWING AND THE PROPERTY OR INFORMATION CONTAINED HEREIN IS THE PROPERTY OF T.H.E. MEDICAL.

E. & O.E.



Installation Drawing Open-Web Steel Joist Mounting & Lateral Bracing

Scale: NTS
July 16, 2007

Dwg Reference
6-9-2b-02

Drawn by
Dan Schwartz